

## Recreational and Surface Water Scheme River, Lake and Stream Water Samples Preliminary Intended Results Notification

These results are provided for guidance and are determined from the examinations undertaken during quality control testing of the samples in the FEPTU laboratory. The enumeration results may show minor variations from the final results, which are calculated from the participants' consensus results, and included in the distribution report.

Distribution: **S121**      Samples: **S121A and S121B**

### Intended results:

| Enumeration                             | S121A<br>Median | S121B<br>Median |
|---|-----------------|-----------------|
| Coliform bacteria in 100mL              | 86              | 0               |
| <i>Escherichia coli</i> in 100mL        | 57              | 0               |
| Enterococci in 100mL                    | 39              | 45              |
| Faecal coliforms in 100mL               | 57              | 0               |
| <i>Clostridium perfringens</i> in 100mL | 66              | 38              |
| <i>Salmonella</i> spp.                  | Detected        | Not detected    |

### Sample contents:

|              |   |
|--------------|---|
| <b>S121A</b> | <i>Klebsiella pneumoniae</i> , <i>E. coli</i> , <i>Enterococcus faecalis</i> , <i>Clostridium perfringens</i> , <i>Salmonella</i> Typhimurium and <i>Staphylococcus saprophyticus</i> |
| <b>S121B</b> | <i>Enterococcus faecalis</i> , <i>Clostridium perfringens</i> and <i>Pediococcus pentosaceus</i>  |

Laboratory identification numbers of participants who did not return results this distribution:

All laboratories returned a result

The distribution report providing details of your individual performance and an analysis of all participants' results will be dispatched within three weeks. The report may also include sample specific comments.

**Repeat samples** are available for up to one year from date of this letter for participants who experienced problems with any of the examination. There is no need to wait until you have received the report to order repeat samples.

The next distribution of Recreational and Surface Water Scheme samples will be Bathing beach and will be dispatched during the week beginning 17 February 2025.

If you require any technical advice or wish to make comments or suggestions relating to the scheme, please contact us.

Validated by:

Divya George  
Healthcare Science  
Specialist

Date: 10 February 2025



0006