

Wastewater Preliminary Intended Results Notification

These results are provided for guidance and are determined from the examinations undertaken during quality control testing of the samples in the FEPTU laboratory. The enumeration results may show minor variations from the final results, which are calculated from the participants' consensus results, and included in the distribution report.

Distribution: **WW001** Samples: **WW001A & WW001B**

Intended results:

Examinations	WW001A Median	WW001B Median
<i>Escherichia coli</i> in 100mL	4.1x10 ⁵	0
Total coliforms in 100mL	4.1x10 ⁵	2.4x10 ²
Enterococci in 100mL	6.0x10 ⁶	5.9x10 ³
<i>Salmonella</i> spp. per L	Not detected	Detected
SARS-CoV-2 per 200mL	Detected	Detected

Sample contents:

WW001A	<i>E. coli</i> , <i>Pantoea agglomerans</i> , <i>Enterococcus faecium</i> , SARS-CoV-2 Omicron B.1.1.529 and <i>Pseudomonas aeruginosa</i>
WW001B	<i>Citrobacter braakii</i> , <i>Enterococcus faecalis</i> , <i>Salmonella</i> Wentworth and SARS-CoV-2 Beta B.1.351

Laboratory identification numbers of participants who did not return results this distribution:

The distribution report providing details of your individual performance and an analysis of all participants' results will be dispatched within three weeks. The report may also include sample specific comments.

Repeat samples are available for up to one year from date of this letter for participants who experienced problems with any of the examination. There is no need to wait until you have received the report to order repeat samples.

The next distribution of Wastewater samples will be dispatched during the week beginning 16 February 2026.

If you require any technical advice or wish to make comments or suggestions relating to the scheme please contact us.

Validated by:	Divya George Healthcare Science Specialist	Date: 1 September 2025
---------------	--	---------------------------